



“Everything in life is vibration.”

– Albert Einstein

Raman Spectroscopy

Optical filters and the value they bring to Raman spectroscopy





“I’m pickin’ up good vibrations.”

– Brian Wilson, the Beach Boys

Abstract:

Raman spectroscopy is a powerful analytical technique that provides valuable insights into the molecular composition and structure of materials.

To achieve accurate and precise Raman measurements, optical filters play a crucial role in enhancing signal-to-noise ratio, and reducing background interference. This white paper explores the fundamentals of optical filters, their types, and their indispensable role in Raman spectroscopy applications. It will also discuss the importance of selecting appropriate optical filters to optimize Raman spectroscopy experiments and enhance the overall analytical performance.

Introduction to Raman spectroscopy

Raman spectroscopy is a non-destructive analytical technique that uses the inelastic scattering of monochromatic light to probe molecular vibrations in a sample. The Raman scattering process yields unique spectral information, allowing researchers to identify chemical compounds, study molecular structures, and monitor chemical reactions.

The technique finds applications in diverse areas, including pharmaceuticals, forensics, environmental science, materials science, geology, and biology, among others. Its benefits are many and varied. For example, it enables the identification of molecules and compounds present in a sample, including organic and inorganic substances, pharmaceuticals, polymers, minerals, and biomolecules. This makes it a valuable tool in analytical chemistry and material science. It is also a non-destructive technique that requires little to no sample preparation, allowing for the investigation of delicate or valuable samples.

In most cases, Raman spectroscopy can be performed with little or no sample contact, reducing the risk of cross-contamination and preserving the sample's integrity. It can also provide real-time or near-real-time results, making it suitable for fast analysis in various applications, including pharmaceutical quality control and process monitoring.

Its versatility allows Raman spectroscopy to be applied to a wide range of samples, including solids, liquids, and gasses, and it can analyze small regions within a sample. This capability is useful in identifying heterogeneous distributions, mapping molecular structures, and studying surface properties. Raman spectroscopy can also be combined with other analytical techniques, such as infrared spectroscopy and mass spectrometry, or microscopy to provide a more comprehensive understanding of a sample's composition and structure. However, Raman signals can be weak or obscured by fluorescence and other background interferences.

Optical Filters in Raman spectroscopy

Optical filters selectively transmit or block specific wavelengths of light based on their spectral characteristics. They can be broadly categorized into two types, edge filters and bandpass filters, though other optics may be required for proper Raman measurements depending on the system being used. Both edge and bandpass filters are used in all Raman spectroscopic instruments, but each serves different purposes. To function as an effective Raman filter, they all must have steep edges so that Raman signals with wavenumbers of 200cm^{-1} can be detected.

2.1 Edge filters

Edge filters, also known as longpass filters, are designed to transmit light with wavelengths longer than a specific cut-on wavelength. In Raman spectroscopy, longpass edge filters are employed to block Rayleigh scattered light, which has the same wavelength as the incident laser light. This rejection of the intense Rayleigh scattering allows for isolation and detection of the much weaker Raman-scattered photons. These are typically used at normal incidence (0deg) to the incoming light path.

2.2 Bandpass filters

Bandpass filters work by passing a specific narrow range of wavelengths and rejecting all other wavelengths outside of that band and are typically used at normal (0deg) incidences. Many modern diode laser sources emit multiple wavelengths or generate less-than-monochromatic light. Bandpass filters are almost always required to act as a laser

“clean-up” filter to limit the spectral output of these lasers, which, if left unchecked, can supersede or veil the Raman signal being measured and create an impure signal.

2.3 Dichroic mirrors

Dichroics are typically used in a microscope-based Raman system at 45 degrees to redirect and reflect the laser output towards the sample under investigation, as well as transmit any signal generated back to a detector. Though these optics can add to the rejection of Rayleigh scattering in their reflection ranges, they do so much less efficiently than edge filters and so should not be relied upon for appropriate signal detection in the absence of an edge filter. The transitional slope of these dichroic mirrors between maximum reflection of the laser and transmission of the expected Raman signal should also be optimized for collection of expected wavenumbers.

2.4 Notch filters

Notch filters are specialized optics that possess a very narrow rejection range with maximal transmission efficiencies for both shorter and longer wavelengths. Used at normal incidence, these can allow for detection of Raman signals in both Stokes and anti-Stokes directions, while also attenuating the incident laser excitation. One application for these filters is in Coherent Anti-Stokes Raman (CARS) imaging.

Factors affecting filter selection:

Several key factors must be considered when selecting optical filters for Raman spectroscopy.

3.1 Laser excitation wavelength

The choice of laser excitation wavelength is fundamental to Raman spectroscopy. The optical filters used must be compatible with the laser's excitation wavelength to effectively block the Rayleigh scattering and transmit the Raman-shifted wavelengths. Depending on the sample being probed, the choice of laser wavelength can lead to a fluorescence signal on top of a Raman signal, giving confusing results. Often, choosing a different wavelength, dual wavelengths, or employing special techniques can reduce or eliminate that fluorescence output.

3.2 Filter's optical properties

The optical filters must have high transmission efficiency for the desired Raman spectral region, provide a steep cut-on to reject the lasers, and transmit low wavenumber Raman output.

3.3 Signal-to-noise ratio (SNR)

An ideal filter should enhance the SNR by reducing background noise. Deep blocking characteristics along with very high transmissions create high value SNR necessary in Raman filter sets.

Sourcing optical filters for Raman spectroscopy

It is important to ask the right questions to ensure that the chosen supplier can effectively meet the filter requirements for your particular Raman spectroscopy application or project.

4.1 Capability

Does your supplier offer the range of optical filters you require, whether you need higher transmission values, steep transitions, or superior blocking to keep out unwanted photons? Are you collecting the proper wavenumbers to efficiently identify your expected samples?

4.2 Experience and expertise

Find out if your supplier has a proven track record of expertise in optical filter design and manufacturing. An experienced supplier will have a deep understanding of the unique challenges and requirements of Raman spectroscopy applications and can even provide tailored solutions to meet your specific needs. They should also demonstrate excellent customer support, technical expertise and value for your money.

4.3 Customization

Raman spectroscopy setups can vary significantly based on the laser excitation wavelength, sample characteristics, and experimental requirements. A good supplier should offer customization options to tailor filters to your specific application, ensuring optimal performance.

4.4 Reputation and customer reviews

Research the supplier's reputation. Look for customer reviews and testimonials to gauge their overall satisfaction with the filters and services provided.

About Chroma Technology

Chroma Technology is a leading manufacturer of Raman filter sets of very durable, and highly precise optical filters using 1st surface hard coat sputter technology for OEM and researchers alike. The company has been supplying solutions for industries ranging from the life sciences and agriculture to manufacturing, inspection, security and aerospace for more than 30 years. Chroma Technology's reputation is built on dedicated customer service, including free technical and applications support.

5.1 Chroma Technology's optical filter solutions for Raman spectroscopy

Chroma Technology offers Raman filters, and sets that are designed for the most common Raman wavelengths. These include: UV (266nm), Visible (488nm, 532nm and 633nm) and NIR (633nm, 785nm and 1064nm).

5.2 Standard Raman filters and sets

Chroma's standard filters offer high transmission for detection of weak signals, as well as broad transmission band, excellent image quality, ultra-steep edges and deep blocking. These are some of the most advanced optical filters available, thanks to Chroma's expertise and experience in precision thin-film design and deposition technology.

Standard Raman filter sets include a narrow-band laser clean-up filter matched with a laser dichroic paired to a long pass emission filter, each with sheer spectral slopes. Together, these filters allow for the detection of even the faintest Raman signals.

5.3 Cost-effective Raman sets

As well as the popular standard Raman filter sets, Chroma also offers a range of cost-effective versions of 785nm and 532nm Raman sets, designed to deliver exceptional performance without breaking the bank. These cost-effective Raman sets boast the same or similar designs as the standard catalog sets, ensuring uncompromised quality and accuracy, but they are tailored to offer slightly relaxed specifications for applications that do not require overengineering, particularly in the flatness of dichroic beamsplitters (DCs) and the steepness of long-pass filters (LPs).

OEM customers in China have placed their trust in these filters, using them in portable and handheld Raman devices for use in security checks, effective detection of explosives and drugs, food safety, and empowering laboratory applications. The ultra-steep filters are widely used by materials science researchers in Raman microscopes.

5.5 Customization capability

If you're looking for something a bit different, or you aren't sure what might best work for your particular application, the expert team at Chroma is here to help. Simply get in touch and let the team know what you are looking to achieve, and if the existing range of products don't quite fit, they will take you through the process of creating something custom that will work.

The process is straightforward. Send us a drawing or let us know what you are looking to achieve. Once the engineering team has all of the specifications, they will design a custom set for your application. Already have something that isn't functioning like you think it should, our team has the breadth of experience to troubleshoot and find out why, without overengineering a solution.

5.6 How can Chroma Technology help me find the right filter for my Raman spectroscopy application?

Get in touch with the team at Chroma today to get a head start on your Raman spectroscopy application without delay.

Phone: +1 800.824.7662

Email: sales@chroma.com

Website: www.chroma.com/request-a-quote

For more information on Chroma Technology's range of optical filters for Raman spectroscopy, visit: [**https://www.chroma.com/raman**](https://www.chroma.com/raman)

**“Ask the right questions and nature
will open the doors to her secrets.”**

– Dr. C. V. Raman, Nobel Prize recipient for
discovery of the Raman Effect

US Corporate Headquarters
10 Imtec Lane
Bellows Falls, Vermont 05101 US
info@chroma.com
Tel: +1-800-824-7662
+1-802-428-2500

Chroma Technology GmbH
Maximilianstrasse 33
D-82140 Olching, Germany
europe@chroma.com
Tel: +49-8142-2847525

Chroma Technology Japan
8F Yokohama Onoecho Building
4-57 Onoecho Naka-ku
Yokohama 231-0015 Japan
Phone: +81 (0) 45 285 1583
Fax: +81 (0) 45 285 1501
japan@chroma.com

Chroma China
Rm 803, Building No.16
Yuanbo Wuli
Xingjin Road, Jimei District
Xiamen City, Fujian Province 361021
Phone: +86-0592-5062089
china@cn.chroma.com

